Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE

Calibration File No.:	01222	Calib. Date:	06/10/2015	Calib. No.:	00034
Certification File No.:	01223	Cert. Date:	06/10/2015	Cert. No.:	00022
Linearity File No.:	01224	Lin. Date:	06/10/2015	Lin. No.:	00022
Solution File No.:	01253	Soln. Date:	09/20/2015	Soln. No.:	00182
Sequential File No.:	01253	File Date:	09/20/2015		

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 04/13/2017
Solution Control Lot: 15D138 Bottle No.: 0005

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:23D		
Control 1 EC	0.100%	13:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:24D		
Control 2 EC	0.100%	13:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:25D		
Control 3 EC	0.100%	13:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: LAROCCA First Name: NICHOLAS MI: M
Signature: N. [illegible] #585 Badge No.: 585
Date: 09/20/2015