

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01604	Calib. Date: 07/29/2019
Certification File No.:	01605	Calib. No.: 00046
Linearity File No.:	01606	Cert. Date: 07/29/2019
Solution File No.:	01626	Cert. No.: 00033
Sequential File No.:	01626	Lin. Date: 07/29/2019
		Lin. No.: 00032
		Soln. Date: 12/07/2019
		Soln. No.: 00259
		File Date: 12/07/2019

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 0769

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:13S		
Control 1 EC	0.099%	18:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:15S		
Control 2 EC	0.096%	18:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:16S		
Control 3 EC	0.096%	18:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: \_\_\_\_\_



Badge No.: 68

Date: 12/07/2019