

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File-No.:	01377	Calib. Date: 09/22/2016
Certification File No.:	01378	Calib. No.: 00038
Linearity File No.:	01379	Cert. Date: 09/22/2016
Solution File No.:	01407	Cert. No.: 00025
Sequential File No.:	01407	Lin. Date: 09/22/2016
		Lin. No.: 00025
		Soln. Date: 02/02/2017
		Soln. No.: 00208
		File Date: 02/02/2017

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %: 0.100%		Expires: 10/28/2018
Solution Control Lot: 16330		Bottle No.: 0590

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:12S		
Control 1 EC	0.101%	07:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:13S		
Control 2 EC	0.101%	07:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	07:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:14S		
Control 3 EC	0.101%	07:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MOHR

First Name: DANNY

MI:

Signature: \_\_\_\_\_



Badge No.: 471

Date: 02/02/2017