

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01548	Calib. Date: 09/26/2018
Certification File No.:	01549	Calib. No.: 00044
Linearity File No.:	01550	Cert. Date: 09/26/2018
Solution File No.:	01574	Cert. No.: 00031
Sequential File No.:	01574	Lin. Date: 09/26/2018
		Lin. No.: 00030
		Soln. Date: 01/15/2019
		Soln. No.: 00242
		File Date: 01/15/2019

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 09/14/2019
Solution Control Lot:	17290		Bottle No.: 0648

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:30S		
Control 1 EC	0.099%	20:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:32S		
Control 2 EC	0.097%	20:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:33S		
Control 3 EC	0.099%	20:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: \_\_\_\_\_



Badge No.: 68

Date: 01/15/2019