

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01579 **Calib. Date:** 03/07/2019 **Calib. No.:** 00045
Certification File No.: 01580 **Cert. Date:** 03/07/2019 **Cert. No.:** 00032
Linearity File No.: 01581 **Lin. Date:** 03/07/2019 **Lin. No.:** 00031
Solution File No.: 01582 **Soln. Date:** 03/07/2019 **Soln. No.:** 00245
Sequential File No.: 01582 **File Date:** 03/07/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWJ S3-0363
Control Solution %: 0.100% **Expires:** 09/14/2019
Solution Control Lot: 17290 **Bottle No.:** 0662

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:16S		
Control 1 EC	0.103%	11:17S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18S		
Control 2 EC	0.099%	11:18S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19S		
Control 3 EC	0.099%	11:20S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number:

DDWAP2-216

BMG

Changed By:

Last Name: GAMBONE

First Name: BRIAN

MI: M

Signature:

TPR. J. B. M. #7029

Badge No.: 7029

Date: 03/07/2019