

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWF-0382 |
| Location: | WEST WINDSOR POLICE | |
| Calibration File No.: | 01548 | Calib. Date: 09/26/2018 |
| Certification File No.: | 01549 | Calib. No.: 00044 |
| Linearity File No.: | 01550 | Cert. Date: 09/26/2018 |
| Solution File No.: | 01561 | Cert. No.: 00031 |
| Sequential File No.: | 01561 | Lin. Date: 09/26/2018 |
| | | Lin. No.: 00030 |
| | | Soln. Date: 11/26/2018 |
| | | Soln. No.: 00239 |
| | | File Date: 11/26/2018 |

| | | | |
|-----------------------|--------|------------------|--------------------------|
| Calibrating Unit: | WET | Model No.: CU-34 | Serial No.: DDWJ S3-0364 |
| Control Solution %: | 0.100% | | Expires: 09/14/2019 |
| Solution Control Lot: | 17290 | | Bottle No.: 0668 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 18:47S | | |
| Control 1 EC | 0.099% | 18:47S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 18:47S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:48S | | |
| Control 2 EC | 0.100% | 18:49S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 18:49S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:49S | | |
| Control 3 EC | 0.099% | 18:50S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 18:50S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:51S | | |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: 

Badge No.: 089

Date: 11/26/2018