

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01471	Calib. Date:	12/04/2017
Certification File No.:	01472	Cert. Date:	12/04/2017
Linearity File No.:	01473	Lin. Date:	12/04/2017
Solution File No.:	01479	Soln. Date:	12/29/2017
Sequential File No.:	01479	File Date:	12/29/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	09/14/2019
Solution Control Lot:	17290			Bottle No.:	0684

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:59S		
Control 1 EC	0.097%	08:00S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00S		
Control 2 EC	0.097%	08:01S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02S		
Control 3 EC	0.097%	08:02S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BROWN

First Name: KYLE

MI: W

Signature: \_\_\_\_\_



Badge No.: 89

Date: 12/29/2017