

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01471 Calib. Date: 12/04/2017 Calib. No.: 00042
Certification File No.: 01472 Cert. Date: 12/04/2017 Cert. No.: 00029
Linearity File No.: 01473 Lin. Date: 12/04/2017 Lin. No.: 00028
Solution File No.: 01491 Soln. Date: 02/17/2018 Soln. No.: 00228
Sequential File No.: 01491 File Date: 02/17/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363
Control Solution %: 0.100% Expires: 09/14/2019
Solution Control Lot: 17290 Bottle No.: 0080

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:23S		
Control 1 EC	0.096%	18:24S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	18:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:25S		
Control 2 EC	0.096%	18:25S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	18:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:26S		
Control 3 EC	0.097%	18:27S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	18:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: PH.  # 585

Badge No.: 585

Date: 02/17/2018