

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01513	Calib. No.: 00043
Certification File No.:	01514	Cert. No.: 00030
Linearity File No.:	01515	Lin. No.: 00029
Solution File No.:	01525	Soln. No.: 00232
Sequential File No.:	01525	File Date: 05/21/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0078

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:01D		
Control 1 EC	0.098%	17:02D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:03D		
Control 2 EC	0.098%	17:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:04D		
Control 3 EC	0.099%	17:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: \_\_\_\_\_

Badge No.: 585

Date: 05/21/2018